



QPS Evaluation Services Inc
 Testing, Certification and Field Evaluation Body
 Accredited in Canada, the USA, and Internationally

File
LR1406

CERTIFICATE OF COMPLIANCE
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Galvanic Applied Sciences Inc.
Address	7000 Fisher Rd. SE Calgary, AB T2H 0W3
Project Number	LR1406-4R1
Product	AccuLase D2 Analyzer with single or Dual Laser Driver
Model Number	<ul style="list-style-type: none"> • Single Laser Driver, Model AccuLase-GPA-D2. • Dual Laser Driver, Model AccuLase-GPA-M-D2
Ratings	Input Voltage: 100 – 240 VAC, 230 watts, 50/60Hz, 1 Phase Hazardous locations Ratings: Class I, Division 2, Groups B, C and D. T3 Tamb: -20°C to + 50°C. Type: 4, 4X
Applicable Standards	CSA C22.2 No. 61010-1:12 UL 61010: 2012 (as a guide) CSA C22.2 No. 213-17 UL 121201
Factory/Manufacturing Location	Same as Applicant

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Dave Adams, P. Eng.
 Manager, Hazardous Locations [Ex Equipment] Department

Signature: 

Date: March 5, 2020