



**QPS Evaluation Services Inc**  
**Testing, Certification and Field Evaluation Body**  
**Accredited in Canada, the USA, and Internationally**

**File**  
**LR1406**

**CERTIFICATE OF COMPLIANCE**  
**(ISO TYPE 3 CERTIFICATION SYSTEM)**

Issued to	Galvanic Applied Sciences Inc.
Address	7000 Fisher Road SE Calgary, AB T2H 0W3
Project Number	LR1406-2R4
Product	Analytes in Water Analyzer
Model Number	AccuColor or AccuSense AccuColor GP or AccuSense GP
Electrical Ratings	100-240 Vac, 50/60 Hz, 310W
Ex Markings	Class I, Division 2, Groups BCD, T3 Ta=-20°C to +60°C
Applicable Standards	ISA 12.12.01:2015 CSA C22.2 No. 213-16 CSA C22.2 No. 61010-1:12 UL 61010-1 3rd ed.
Factory/Manufacturing Location	Galvanic Applied Sciences USA 101 Billerica Avenue, Building 5, Suite 104 North Billerica, MA 01862 USA
Special Conditions of Use	N/A

**Statement of Compliance:** The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.

**IMPORTANT NOTE:** Certification will be revoked if compliance to the latest versions of the standard(s) is not maintained and/or if the product/equipment is modified after certification is granted without prior written consent by QPS.



Issued By: Rob Kohuch P.Eng.  
Senior Engineer - Hazardous Locations [Ex Products]

Signature: *Rob Kohuch*

Date: October 2, 2023



The SCC and IAS Accreditation Symbols are official symbols of the respective accreditation bodies, used under license.  
 81 Kelfield St., Units 8, Toronto, ON M9W 5A3 Tel: 416-241-8857; Fax: 416-241-0682  
 www.qps.ca

