



**QPS Evaluation Services Inc**  
 Testing, Certification and Field Evaluation Body  
 Accredited in Canada, the USA, and Internationally

File <b>LR1406</b>
-----------------------

**CERTIFICATE OF COMPLIANCE**  
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Galvanic Applied Sciences Inc.
Address	7000 Fisher Road SE Calgary, Alberta T2H 0W3
Project Number	LR1406-8
Product	Gas Chromatograph
Model Number	AccuChrome-D2-S AccuChrome-D2-C6 AccuChrome-D2-C9
Ratings	Class I, Division 2, Groups BCD T3 T <sub>amb</sub> = -20°C to +60°C 100-240 VAC, 1 Phase, 50/60 Hz, 150 watts OR 24 VDC, 150 watts
Applicable Standards	CSA C22.2 No 61010-1-12 UL 61010-1 3 <sup>rd</sup> ed CSA C22.2 No. 213: 2017 UL 121201 9 <sup>th</sup> ed
Factory/Manufacturing Location	Galvanic Applied Sciences Inc. 7000 Fisher Road SE, Calgary, Alberta T2H 0W3

**Statement of Compliance:** The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Rob Kohuch, P. Eng.  
 Hazardous Locations [Ex Equipment] Department

Signature: *Rob Kohuch*

Date: November 11, 2022